

LN2 Refill Without Samples Inside Instructions

NEED HELP? CONTACT CRYOPORT SYSTEMS' CLIENT CARE AT (949) 470-2305 OR EMAIL CS@CRYOPORT.COM

## Step 1










Wear the appropriate Personal Protective Equipment (PPE) when handling cryogenic shipments.

#### Step 2




Contact Cryoport Systems' Client Care to access the In-Field Recharging Verification Form and complete Section A.


#### Step 3



To lift the metal handle, see Step 4. Once the handle is up, hold it, place your foot at the bottom of the metal panel, tilt the shipping system back, and use the integrated wheels to move it to the desired location.

### Step 4






To unlock the metal handle, use one hand to slide the lever to the right and the other to lift or lower the handle.

NOTE: If the metal handle is bent or broken, use the nested handles on the sides to move the shipping system.

WARNING: DO NOT CARRY THE SHIPPING SYSTEM BY THE HANDLE.


## Step 5



Use scissors to remove zip ties from both steel latches.

NOTE: If a serialized zip tie is present, record the serialized number on all necessary documentation.

## Step 6



To unlock the latches, push the button located behind the latch handle and lift. To disconnect the top, pull the latch away from the hooks.



LN2 Refill Without Samples Inside Instructions

NEED HELP? CONTACT CRYOPORT SYSTEMS' CLIENT CARE AT (949) 470-2305 OR EMAIL CS@CRYOPORT.COM

## Step 7



To access the dewar, open the lid and cut off the zip tie on the hinged cap with scissors.

NOTE: If a serialized zip tie is present, record the serialized number on all necessary documentation.


#### Step 8



#### For SET-1000 Safepak<sup>®</sup> Soft System 1800 ONLY

If using a Safepak® Soft System 1800, DO NOT complete Steps 9, 11-14, and 16. Contact Cryoport Client Care to confirm the amount of LN2 you should add for your specific use case. DO NOT disconnect the dewar from the enclosure to pour out excess LN2.


### Step 9



Disconnect the dewar from the enclosure by unlatching the carabiner behind the dewar. Use the metal handle to remove the dewar from the enclosure and place it on a flat surface.

NOTE: Follow local safety regulations associated with proper lifting.

## Step 10



To remove the vapor plug, pull up on the circular handle located in the center. Place the vapor plug aside with the fleece lining facing up.


NOTE: If the dewar does not emit vapor after the plug is removed, please call Cryoport Systems' Client Care at (949) 470-2305.

## Step 11



Remove the absorbent pad from inside of the dewar and place it on the enclosure lid.

## Step 12




Slowly pour Liquid Nitrogen into the opening. Fill the dewar up to the bottom of the neck tube. Reinsert the vapor plug. Close the hinged cap by pulling down on the lid. Let the dewar stabilize in a static position for at least 60 minutes.

Cryoport Systems

LN2 Refill Without Samples Inside Instructions

NEED HELP? CONTACT CRYOPORT SYSTEMS' CLIENT CARE AT (949) 470-2305 OR EMAIL CS@CRYOPORT.COM

#### Step 13



Once the dewar has stabilized, open the hinged cap and pull up on the circular handle in the center to remove the vapor plug. Set the vapor plug aside, with fleece lining face up.


#### Step 14



Safely pour the excess LN2 into a compatible LN2 container.

NOTE: To avoid pinch points lift the dewar properly (as shown) and not by the handle. This process may take two people to complete. Follow local safety regulations associated with proper lifting.


#### Step 15



Reinsert the absorbent pad and the vapor plug. Close the hinged cap and secure it with a zip tie. Do not overtighten the zip tie as it will cause breakage.

NOTE: If a serialized zip tie is present, record the serialized number on all necessary documentation.

## Step 16



Carefully place the dewar back into the enclosure and reconnect the dewar by latching the carabiner behind the


NOTE: Follow local safety regulations associated with proper lifting.

## Step 17



Complete Section C and D of the In-Field Recharging Verification Form.

## Step 18




Close the lid and align it with the base. After aligning the enclosure, lift the latch handles and connect them to the hooks. Then, pull the latch handle down towards the enclosure until you hear a click to secure the latch. Complete this step for both latches.

Cryoport Systems

LN2 Refill Without Samples Inside Instructions

NEED HELP? CONTACT CRYOPORT SYSTEMS' CLIENT CARE AT (949) 470-2305 OR EMAIL CS@CRYOPORT.COM


### Step 19



Insert the zip tie from left to right through each metal latch hole, wrap it around each latch, tighten it, and secure the lid.

NOTE: If a serialized zip tie is present, record the serialized number on all necessary documentation.

#### Step 20



Place the shipping system on flat ground. See Step 4 to lift the metal handle with your hands, then place your foot at the bottom of the metal panel, tilt the shipping system back, and use the integrated wheels to move it to the desired location.